

A novel soft X-ray PGM for elliptical dipole radiation

Torsten Kachel^{a*}, Frank Eggenstein^b and Rolf Follath^c

^aInstitute for Methods and Instrumentation for Synchrotron Radiation Research, Helmholtz-Zentrum Berlin für Materialien und Energie, Albert-Einstein-Strasse 15, Berlin, 12489, Germany

^bInstitute for Nanometer Optics and Technology, Helmholtz-Zentrum Berlin für Materialien und Energie, Albert-Einstein-Strasse 15, Berlin, 12489, Germany

^cBeamline Optics Group, Paul Scherrer Institute, Villigen, 5232, Switzerland

Correspondence email: torsten.kachel@helmholtz-berlin.de

Synopsis The paper describes the utilization of elliptical dipole radiation in a collimated plane grating monochromator at BESSY II.

Abstract We describe a new but yet well proven way of making elliptically polarized dipole radiation from the BESSY II storage ring applicable to the SX700-type collimated plane grating monochromator PM3. We show that due to the limited vertical acceptance of the grating a simple use of vertical apertures is not possible in this case. Rather, deflecting the beam up- or downwards by rotating the vertically collimating toroidal mirror M1 around the light axis leads to an excellent performance. The resulting detune of the photon energy can be taken into account by a readjustment of the monochromator internal plane mirror M2. The energy resolution of the beamline is not affected by the non-zero “roll” of the collimating mirror.

1. Introduction

For more than two decades the use of circularly polarized soft X-rays has developed a major tool of studying magnetic thin films and surfaces. Namely, the application of X-ray magnetic circular dichroism (XMCD) (Schütz et al., 1987; Chen et al., 1990; Baumgarten et al., 1990; for an extensive overview see e.g. Stöhr & Siegmann) offers a wide field of experimental applications. Whenever the orientation between the k-vector of the impinging elliptically polarized soft X-rays and the sample magnetization vector is changed a variation of the absorption coefficient is observed. Today this fact is well known and understood (Thole et al., 1985). The XMCD effect can be seen in numerous

experimental applications such as e.g. photoemission spectroscopy, total electron yield spectroscopy, X-ray absorption, transmission and reflectivity measurements.

Therefore, the need for beamlines delivering elliptically polarized synchrotron radiation (SR) and - as a consequence - the number of such beamlines has increased significantly. An important improvement was the extension of the photon energy range from low energy ($h\nu \leq 25$ eV) normal incidence monochromators (NIM) (Schäfers et al., 1986) to grazing incidence plane grating (PGM) or spherical grating (SGM) monochromators and the implication of crossed (Bahrdt et al., 1992) or elliptical (Sasaki et al., 1992) undulators delivering about 100 times more photon flux than dipole sources.

Along with these improvements considerable progress has been made in soft X-ray beamline design. This holds predominantly concerning properties like brilliance, energy tuneability, energy resolution and focal spot size. In particular, and important in the context of this paper, the development of the so called “collimated PGM” (cPGM) (Follath & Senf, 1997) has been established a significant design improvement. A photon energy resolution down to the diffraction limit and tuneability of the c_{ff} -value that allows for easy energy calibration and higher order suppression are only few of its features.

It is understood that the combination of both, using elliptically polarized dipole radiation and applying a cPGM is not for free unless additional effort is made. However, here we will outline that a simple rotation of the collimation mirror can combine both objectives.

2. PM3 beamline layout

Figure 1 Beamline layout of PM3 at BESSY II. The dipole source is on the left, the exit slit on the right. The toroidal mirror M1 serves for vertical (sagittal) collimation of the synchrotron light.

The beamline layout follows the collimated design that was implemented in several beamlines at BESSY II before (Follath 2001). Figure 1 shows the design of PM3. It consists of four optical elements.

The toroidal mirror M1, at 13000 mm from the source, collimates the divergent beam vertically (sagittally). Horizontally (meridionally) it focusses the beam with unit magnification at 26000 mm. M2 and the grating are plane elements in the standard Zeiss SX700 configuration. This is the same, partly upgraded, monochromator that has been in operation at BESSY I as “SX700-III” (Petersen et al., 1993). The grating is located at 17000 mm from the source. All experimental data and calculations in this paper are related to the high resolution 1221 1/mm grating.

Because of its restricted length of 120 mm and typical incidence angles α of less than 5° the grating has the smallest vertical acceptance of all four optical elements and defines therefore the vertical acceptance of the beamline.

In figure 2 we present calculated values for the incidence angle α as well as the vertical acceptance in mm and in mrad of the 120 mm long grating. The values are parameterized for typical c_{ff} values: $c_{ff} = 2.25$ (standard operation), $c_{ff} = 1.4$ (higher order suppression mode), and $c_{ff} = 5$ (high resolution mode). We note that the acceptance in mm and mrad are connected through the 13000 mm entrance arm length of M1.

Figure 2 Vertical acceptance of the BESSY II PM3 grating for varying c_{ff} values.

3. Source characteristics of BESSY II dipoles

It is well-known that SR from dipole magnets is polarized (Schwinger, 1949). For a further discussion we define the off-plane viewing angle ψ and the angle of acceptance $\Delta\psi$. Both quantities might be defined by vertical apertures. The position of the center of acceptance defines ψ whereas the size is a measure for $\Delta\psi$. At $\psi = 0$, i.e. within the storage ring plane the polarization is linear with the polarization vector lying in the horizontal plane. At finite ψ the polarization is right or left handed elliptical for negative and positive viewing angles, respectively.

For given storage ring parameters, i.e. fixed electron energy and magnetic dipole field, the vertical width of the emission cone is a function of the photon energy only. For the BESSY II case and selected photon energies the situation appears as shown in figure 3. The curves have been calculated with the REFLEC software (Schäfers & Krumrey, 1996). All panels show vertical distributions for a set of different photon energies which are near the lowest and highest accessible photon energies of the PM3 and an intermediate one of interest (Fe 2p absorption edge).

Figure 3 Calculated flux (top panel), ellipticity (center), and figure of merit (FoM, bottom) for $h\nu = 50, 700$, and 2000 eV from BESSY II dipoles. The bars in the bottom panel indicate the vertical acceptance of the PM3 grating at the color coded photon energies in standard operation ($c_{ff} = 2.25$).

In the calculation we applied vertical acceptance angles $\Delta\psi$ that are determined by the vertical acceptance of the grating in standard operation ($c_{ff} = 2.25$). The horizontal beamline acceptance is determined by the 980 mm optical length of M1 which refers to 3.42 mrad constantly. In the top panel we present the photon flux I . We notice that the FWHMs of the vertical emission cones range from

about 0.6 to 2.5 mrad. Likewise behaves the circular polarization S_3 (which should more precisely be called the Stokes-Poincaré parameter) in the center panel. However, rather than flux I and circular polarization S_3 the quantity of interest for optimal experimental conditions is the figure of merit FoM (Petersen et al., 1993) defined as

$$\text{FoM} = S_3^2 * I \quad (1)$$

The FoMs are depicted in the bottom panel. We observe maxima at $\psi \approx 0.2$ (2000 eV), 0.35 (700 eV) and 0.85 mrad (50 eV). As stated above these values are too large to be accepted by the grating when using the cPGM standard alignment (see figure 2). This is illustrated by the colored bars indicating the photon energy dependent acceptance of the grating. In contrast, the collimation mirror M1 with a vertical optical active width of 40 mm (= 3.1 mrad) covers all necessary viewing angles ψ . From the S-shaped ellipticity curves in the center panel of figure 3 it is evident that in the case of larger viewing angles errors in ψ have little effect on S_3 . In contrast, for π -light the setting of ψ is much more critical. In other words it is hardly possible to achieve π -polarization precisely.

By comparison of figures 2 and 3 it becomes obvious that for $c_{\text{ff}} = 2.25$ we find an incidental very good matching between the required (fig. 2) and naturally given (fig. 3) beamline acceptance ranging from 0.2 to 1 mrad. Therefore, the application of vertical entrance apertures is not required. For higher c_{ff} values the acceptance becomes smaller and, thus, it is always possible to select radiation of high ellipticity. Only in the higher order suppression mode $c_{\text{ff}} = 1.4$ where the acceptance ranges from 0.4 to 2 mrad additional apertures are needed to select elliptically polarized SR with high FoM. Obviously, the same statement holds true when operating the monochromator in outside (negative) diffraction order, where the acceptance is even larger than for $c_{\text{ff}} = 1.4$.

4. Changing the viewing angle ψ by rotation of the collimation mirror

Up to now we have mainly discussed the acceptance $\Delta\psi$. However, the main issue of the setup is the viewing angle ψ . Figure 3 shows that SR of high FoM cannot be accepted by the grating when using the standard alignment. In the past years different principles have been applied to steer the desired part of the SR cone onto the grating, namely employing additional vertically deflecting mirrors (Petersen et al., 1993) or implementing a vertical “bump” to the electron beam within the dipole, i.e. steering the “storage ring plane” (Hunter Dunn et al., 2004; Raabe et al., 2008). We note that the latter principle of steering the electron beam has been tested but found to be not applicable at the third generation storage ring BESSY II (Kachel & Feikes, 2000). Instead in the PM3 we used the roll of M1, i.e. the rotation of M1 around the axis of the incoming light, to steer the required part of the synchrotron beam into the beamline. This is an elegant way to avoid additional mirrors. We refer to this degree of freedom as R_z , denoting the rotation around z (= light axis) of the collimation mirror.

Changing R_z leads to an up- or downward reflection of the SR. It can be tuned such that the maximum FoM lies in the grating center. For all day operation we use seven predefined rotation angles: one for linearly polarized and three for left and right handed elliptical polarization, respectively. Table 1 gives an overview of the related parameters. The fact that the nominal R_z values are not fully symmetric is caused by cross-talking of the actuators for the different degrees of freedom. The asymmetry in the deflection angles arises because the grating center is not perfectly hit under all conditions. It is seen that in our case a free R_z -rotation of $\pm 2^\circ$ is sufficient. We note that for finite R_z also a slight correction of the horizontal deflection is required.

Table 1 M1 rotation around R_z in seven steps.

Naming	Polarization	R_z (μrad)	M2 offset angle($^\circ$)	Energy range (eV)
+1.2	σ^+	32279	0.06150	< 100
+0.8		21862	0.04100	100 – 1000
+0.4		11445	0.02060	> 1000
Pi	π	0	0.00000	full range
-0.4	σ^-	-9388	-0.02050	> 1000
-0.8		-19805	-0.04114	100 – 1000
-1.2		-30222	-0.06190	<100

The column “M2 offset angle” denotes half the vertical beam deflection behind M1.

A side effect of the roll variation of M1 is the change of the incident angle on M2 and the grating that leads to a detuning of the photon energies. This is compensated by a readjustment of the plane mirror M2 rotation. As a result, the photon energy at the sample position stays unaltered after a polarization change. As the amount of M2-rotation depends only on the Roll-angle of M1 it is independent of the photon energy. The offset angles for M2 have been calculated and experimentally determined for each polarization setting using N_2 gas absorption spectra.

The energy resolution of the beamline has been measured by gas absorption spectroscopy. The ion yield of the He 2,-1₄ peak in the Rydberg series gives a total line width of 1.9 meV at about 64 eV photon energy. This corresponds to an energy resolution

$$R = E/\Delta E = 34000 \text{ @ } 64 \text{ eV} \quad (2)$$

in an optimal case using $c_{ff} = 5$. Concerning our new principle of polarization tuning it might rather be relevant to show how the change of R_z affects the resolving power of the beamline. This is due to the fact that a detuning of the M1 roll might well be interpreted as an on-purpose “misalignment” of a critical optical beamline element. The experiment, however, gave an unambiguous prove of a stable high performance under the conditions listed in table 1. In figure 4 we present N 1s ion yield spectra of the N₂ absorption obtained with linearly and elliptically polarized SR from PM3. We conclude that no obvious degradation of the energy resolution is observed in the case of finite roll R_z for elliptically polarized SR.

Figure 4 Ion yield at the N 1s absorption line of N₂. Black: σ^+ -polarization (with vertical offset for better visibility), red: π -polarization. Circles: experiment, lines: least-squares fits of six Voigt peak profiles.

Under typical operation conditions the beamline delivers an experimentally determined ellipticity of 92% at the Fe L₃ edge ($h\nu = 707$ eV). The values are similar for the other transition metal 2p absorption edges. In contrast, photon energies below about 100 eV are rarely used. This is due to the fact that in this energy range only few absorption edges of interest in XMCD work exist and that the degree of ellipticity might be easily obscured by depolarization effects of the beamline (Bahrtdt et al., 2010). But it has been shown that even at the Pt 4f absorption edge at about 72 eV PM3 delivers high degree of circular polarization and precision of XMCD asymmetries much better than 0.1% (Honolka et al., 2009). The highest photon energy used so far with elliptical polarization was 1853 eV for resonant excitation of the Si 1s edge. In this experiment induced magnetic moments in Heusler-like Fe₃Si were studied (Antoniak et al., 2012). Further prominent publications from the PM3 beamline can be found in additional references (Valencia et al., 2011; Radu et al., 2012; Antoniak et al., 2011; Mishra et al., 2009; Sanyal et al., 2010).

5. Summary

We have shown that a state-of-the-art high resolution collimated plane grating monochromator can exploit elliptically polarized dipole radiation from a third generation storage ring without additional mirrors or electron beam steering. A simple rotation R_z of the collimation mirror (M1) around the axis of the incoming SR beam is sufficient to direct the elliptically polarized part of the dipole emission cone onto the grating. The resulting energy shift can be compensated by a R_z depending offset in the plane mirror (M2) rotation angle. No degradation of the energy resolution caused by the steering with M1 could be observed.

Acknowledgements We gratefully acknowledge the contributions of our former colleagues H. Gundlach, M. Willmann and W. Gudat. We thank F. Schäfers and F. Senf for careful reading of the manuscript.

References

- Antoniak, C., Gruner, M. E., Spasova, M., Trunova, A. V., Römer, F. M., Warland, A., Krumme, B., Fauth, K., Sun, S., Entel, P., Farle, M. & Wende, H. (2011). *Nature Comm.* **2**, 528.
- Antoniak, C., Herper, H. C., Zhang, Y. N., Warland, A., Kachel, T., Stromberg, F., Krumme, B., Weis, C., Fauth, K., Keune, W., Entel, P., Wu, R. Q., Lindner, J. & Wende, H. (2012). *Phys. Rev. B* **85**, 214432.
- Bahrtdt, J., Gaupp, A., Gudat, W., Mast, M., Molter, K., Peatman, W. B., Scheer, M., Schroeter, Th. & Wang, Ch. (1992). *Rev. Sci. Instrum.* **63**, 339-342.
- Bahrtdt, J., Follath, R., Frentrop, W., Gaupp, A. & Scheer, M. (2010). *AIP Conf. Proc.* **1234**, 335.
- Baumgarten, L., Schneider, C. M., Petersen, H., Schäfers, F., & Kirschner, J. (1990). *Phys. Rev. Lett.* **65**, 492.
- Follath, R. & Senf, F. (1997). *Nucl. Instr. Methods Phys. Res.* **A390**, 388-394.
- Follath, R. (2001). *Nucl. Instr. Methods Phys. Res.* **A467-468**, 418-425.
- Honolka, J., Lee, T. Y., Kuhnke, K., Enders, A., Skomski, R., Bornemann, S., Mankovsky, S., Minár, J., Staunton, J., Ebert, H., Hessler, M., Fauth, K., Schütz, G., Buchsbaum, A., Schmid, M., Varga, P., & Kern, K. (2009). *Phys. Rev. Lett.* **102**, 067207.
- Hunter Dunn, J., Hahlin, A., Karis, O., Arvanitis, D., LeBlanc, G., Andersson, Å. & Lindgren, L.-J. (2004). *AIP Conf. Proc.* **705**, 65.
- Kachel, T. & Feikes, J. (2000). Unpublished results.
- Mishra, S. K., Radu, F., Dürr, H. A. & Eberhardt, W. (2009). *Phys. Rev. Lett.* **102**, 177208.
- Petersen, H., Willmann, M., Schäfers, F. & Gudat, W. (1993). *Nucl. Instr. and Meth. Phys. Res.* **A333** (2-3), 594-598.
- Raabe, J., Tzvetkov, G., Flechsig, U., Böge, M., Jaggi, A., Sarafimov, B., Vernooij, M. G. C., Huthwelker, T., Ade, H., Kilcoyne, D., Tyliczszak, T., Fink, R. H., & Quitmann, C. (2008), *Rev. Sci. Instrum.* **79**, 113704.
- Radu, F., Abrudan, R., Radu, I., Schmitz, D. & Zabel, H. (2012). *Nature Comm.* **3**, 715.
- Sasaki, S., Miyata, K. & Takada, T. (1992). *Jpn. J. Appl. Phys.* **31**, L1794 –L1796.
- Sanyal, B., Antoniak, C., Burkert, T., Krumme, B., Warland, A., Stromberg, F., Praetorius, C., Fauth, K., Wende, H. & Eriksson, O. (2010). *Phys. Rev. Lett.* **104**, 156402.

- Schäfers, F., Peatman, W., Evers, A., Heckenkamp, Ch., Schönhense, G. & Heinzmann, U. (1996). *Rev. Sci. Instrum.* **57**, 1032.
- Schäfers, F. & Krumrey, M. (1996). Technischer Bericht 201, Berliner Elektronenspeicherring-Gesellschaft für Synchrotronstrahlung.
- Schwinger, J. (1949). *Phys. Rev.* **75**, 1912.
- Stöhr, J. & Siegmann H.C. (2006). *Magnetism: From Fundamentals to Nanoscale Dynamics*. Springer
- Valencia, S., Crassous, A., Bocher, L., Garcia, V., Moya, X., Cherifi, R. O., Deranlot, C., Bouzehouane, K., Fusil, S., Zobelli, A., Gloter, A., Mathur, N. D., Gaupp, A., Abrudan, R., Radu, F., Barthélémy, A., & Bibes, M. (2011). *Nature Mater.* **10**, 753-758